

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  04D1084977	<b>(X3) Date Survey Completed</b>  05/24/2022
<b>Name of Provider or Supplier</b>  Healthcare Express	<b>Street Address, City, State</b>  1509 W Loop 281, Longview, TX	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	. An onsite survey conducted 05/24/2022 found the laboratory in compliance with 42 CFR Part 493, Requirements for Laboratories. .
<b>D2007</b>	<p><b>TESTING OF PROFICIENCY TESTING SAMPLES</b> CFR(s): 493.801(b)(1)</p> <p>The samples must be examined or tested with the laboratory's regular patient workload by personnel who routinely perform the testing in the laboratory, using the laboratory's routine methods</p> <p>This STANDARD is not met as evidenced by: . Based on a review of proficiency testing records, review of the Centers of Medicare and Medicaid Services (CMS) form 209, and confirmed in interview, the laboratory failed to rotate proficiency testing samples through testing persons for nine of fourteen proficiency testing events reviewed from 2020 to 2022 The findings include: 1. Review of the CMS209 personnel form had four testing personnel available for testing in 2020 and six testing personnel available for testing in 2021 for the specialties of hematology and chemistry. 2. Review of the American Proficiency Institute (API) testing records from 2020 and 2022 has the following nine of fourteen events performed exclusively by testing personnel 2. 2020: Chemistry Event 2: TP2 Hematology Event 2: TP2 Chemistry Event 3: TP2 Hematology Event 3: TP2 2021: Chemistry Event 1: TP2 Hematology Event 1: TP2 Chemistry Event 2: TP2 Hematology Event 2: TP2 Hematology Event 3: TP2 3. In an interview on 5/24/2022 at 08:20 hours, in the office, testing personnel 1, as designated by the CMS209, confirmed that the laboratory was not rotating proficiency testing through the testing persons. .</p>