

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 04D2065828	(X3) Date Survey Completed 03/20/2018
Name of Provider or Supplier Sills Dermatology	Street Address, City, State 1003 Windover Rd, Jonesboro, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Through observation and interview it was determined that the laboratory failed to assure that supplies were not used when they have exceeded their expiration date. Findings follow: A. During a tour of the laboratory on 3/20/18 at approximately 1545 one of one bottle of Potassium Hydroxide Solution Lot # 1578-00 with an expiration date of September 2015 was observed in a cabinet next to cover slips and slides. B. In an interview on 3/20/18 at approximately 1545 the office manager identified as number one on the "Entrance and/or Exit Conference Attendance Record" verified that the bottle was available for use and had exceeded the expiration date.</p>