

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  04D2144056	<b>(X3) Date Survey Completed</b>  09/27/2024
<b>Name of Provider or Supplier</b>  Pathnet Llc	<b>Street Address, City, State</b>  5100 Talley Road, Ste 300, Little Rock, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Through observation and interview with laboratory staff it was determined that the laboratory had supplies available for use after their expiration date. Findings follow: A) During a tour of the laboratory on 9/27/24 at 3:18 pm, one (of seven) opened container of Methanol (StatLab, lot 163901, 40380-1, expiration date 9/21/24) was observed in the laboratory, available for use beyond the expiration date. B) During a tour of the laboratory on 9/27/24 at 3:20 pm, one (of one) unopened container of 37% Formaldehyde (StatLab, lot 121812, 28530-1, expiration date 5/31/23) was observed in the laboratory, available for use beyond the expiration date. C) In an interview on 9/27/24 at 3:21 pm the general supervisor, confirmed that the items, identified above, had exceeded the expiration date and were available for use.</p>