

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 04D2148247	(X3) Date Survey Completed 07/15/2026
Name of Provider or Supplier Premier Gastroenterology Associates, PLLC	Street Address, City, State 10915 Rodney Parham, Suite B, Little Rock, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d) (d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality. This STANDARD is not met as evidenced by: Based on observation and interview with laboratory staff, the laboratory had supplies available for use after their expiration date. Findings follow: A) During a tour of the laboratory on 7/15/26 at 9:08am, one (of one) containers of "Blue Tissue Marking Dye" (Statlab, lot: 167606, SL662BL-2, expiration date 2025-04-30) was observed in the laboratory, available for use beyond the expiration date. B) During a tour of the laboratory on 7/15/26 at 9:08am, one (of one) containers of "Black Tissue Marking Dye" (Statlab, lot: 170347, SL662BK-2, expiration date 2025-04-30) was observed in the laboratory, available for use beyond the expiration date. C) In an interview on 7/15/26 at 9:14am laboratory personnel #1 (as identified on entrance/exit conference sign in sheet) confirmed that the items, identified above, had exceeded their expiration date and were available for use in the laboratory.