

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 04D2166077	(X3) Date Survey Completed 11/18/2025
Name of Provider or Supplier Hull Dermatology	Street Address, City, State 9521 Sunset Dr, Bella Vista, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5413	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b)</p> <p>(b) The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (b)(1) Water quality. (b)(2) Temperature. (b)(3) Humidity. (b)(4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.</p> <p>This STANDARD is not met as evidenced by: Based upon observation, review of Mohs surgery log, review of cryostat temperature records and interview with laboratory staff the laboratory failed to document proper cryostat operating temperatures for 550 Mohs surgery procedures performed in 2025. Findings follow: A) During a tour of the laboratory on 11/18/25 at 1:15 p.m. a cryostat with an manufacturer defined acceptable temperature range of -21 degrees C. to -35 degrees C. was observed in the tissue processing area. B) In an interview on 11/18/25 at 01:15 p.m., the laboratory staff member (identified on the entrance conference attendance record) stated that Mohs surgeries are performed on Mondays in the clinic and that 550 Mohs surgery procedures were performed in 2025. C) Review of the cryostat temperature log revealed that no operating temperature was recorded on any date in 2025. D) In an interview on 11/18/25 at 03:00 p.m. the laboratory staff member (identified on the entrance conference attendance record) confirmed that Mohs surgeries were performed in 2025 and the cryostat was used to prepare slides used for diagnosis on those surgeries, and there was no documentation of the proper cryostat operating temperatures in 2025.</p>
D5417	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT

CFR(s): 493.1252(d)

(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.

This STANDARD is not met as evidenced by:

Based on observation and interview with laboratory staff, the laboratory had supplies available for use after their expiration date. Findings follow: A) During a tour of the laboratory on 11/18/25 at 2:45 pm, one (of 1) bottle of Toluidine Blue O, 1% lot# 0027-00, expiration date 2021-08-31 was observed in the laboratory cabinet, available for use beyond the expiration date. C) In an interview on 11/18/25 at 3:00 pm the laboratory staff member (identified on the entrance conference attendance record) confirmed that the item, identified above, had exceeded the expiration date and was available for use in the laboratory.