

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 04D2190122	(X3) Date Survey Completed 05/13/2025
Name of Provider or Supplier Conway Cardiovascular Surgery Center	Street Address, City, State 650 United Drive Suite 120, Conway, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6053	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9)</p> <p>(b)(9) Evaluating and documenting the performance of individuals responsible for moderate complexity testing at least semiannually during the first year the individual tests patient specimens.</p> <p>This STANDARD is not met as evidenced by: . Based on review of the laboratory policy manual, personnel records, lack of documentation, and interview with staff, it was determined that the technical consultant failed to document personnel competency for one of 5 personnel identified on the CMS form 209. Survey findings follow: A. A review of personnel records for testing personnel revealed that the technical consultant failed to perform an semiannual competency evaluation for testing personnel #2 (as listed on form CMS 209) who was hired July 2024. B. Upon request, the laboratory could not provide competency evaluations for the personnel identified above. C. In an interview on 5/13 /24 at 10:58 a.m., the testing personnel #1 confirmed that a six-month competency evaluation had not been performed on testing personnel #2 (as listed on form CMS 209).</p>