

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 04D2298060	(X3) Date Survey Completed 10/03/2024
Name of Provider or Supplier Healthstar Diagnostic Laboratory	Street Address, City, State 1661 Airport Rd, Ste A, Hot Springs, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Observation and interview with laboratory staff determined that the laboratory had supplies available for use after their expiration date. Findings follow: A) During a tour of the laboratory on 10/3/2024 at 1:50pm, one plastic bag containing 20 Sodium Flouride (Blue Top) blood collection tubes lot 3289129, expiration date 2024-07-31 was observed in the phlebotomy area, available for use beyond the expiraton date. B) In an interview on 10/3/24 at 2:00 pm the laboratory personnel (# 2 on form CMS 209), confirmed that the items, identified above, had exceeded the expiration date and were available for use.</p>