

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  05D0544111	<b>(X3) Date Survey Completed</b>  01/07/2021
<b>Name of Provider or Supplier</b>  James Richards, Md	<b>Street Address, City, State</b>  6222 W Manchester Ave, Ste A, Los Angeles, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2087</b>	<p><b>ROUTINE CHEMISTRY</b> CFR(s): 493.841(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on the surveyor's review of the laboratory's records for evaluation of proficiency testing performance from CMS, laboratory proficiency reporting from API, and an interview with laboratory technical consultant (TC) on 1/7/21 between 2:30 p.m. . and 4:00 p.m., it was determined that there were several unacceptable (below 80 %) proficiency testing (PT) results for Chloride (Cl) for Cycle Q3 (2019). Findings include: 1. On 1/7/2021, an inspection was conducted between 2:30 p.m. and 4:00 p. m. 2. During a review of the laboratory documentation from API (agency providing the proficiency specimens), it was noted at approximately 3:30 p.m. that there were several unacceptable PT results for cycle Q3 Chloride on the 2019 report. The laboratory utilizes the Envoy 500 instrument for general chemistry testing The TC recognized these atypical results. 3. The findings and acceptable ranges were as follows: Analyte: Chloride (5/5 Unacceptable) Sample Actual Result Expected Result (range) CH11 101 88-99 CH12 103 91-102 CH13 115 100-112 CH14 116 103-115 CH15 122 109-121 4. The TC affirmed the unacceptable results listed above.</p>