

| | | |
|--|---|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 05D0546234 | (X3) Date Survey Completed 04/03/2018 |
| Name of Provider or Supplier Scrc Medical Group Pc | Street Address, City, State 450 N Roxbury Dr Ste 101, Beverly Hills, CA | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|--|
| D2098 | <p>ENDOCRINOLOGY CFR(s): 493.843(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review second quarter (Q2-2016) of the American Proficiency Institute (API) proficiency testing records, random patient sampling test results, and interview with the testing personnel, it was determined that the laboratory failed to attain a score of at least 80 percent of acceptable responses for Progesterone analyte. The findings included: a. Q2-2016, API reported an unsatisfactory score of 0% for Progesterone test. b. For four (4) out of sixteen (16) random patient sampling test results reviewed covering period from 1/3/2016 to 2/9/2018, four (4) patients have Progesterone test analyzed and reported during the approximate time the laboratory received the unsatisfactory proficiency testing score. c. The testing personnel affirmed (4/3/2018, 1400) that the laboratory received the above unsatisfactory proficiency testing score of 0% for Progesterone test.</p> |