

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 05D0546491	(X3) Date Survey Completed 04/15/2025
Name of Provider or Supplier Comprehensive Urology Medical Group	Street Address, City, State 8631 W 3rd St Ste 1115e, Los Angeles, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A proficiency testing desk review survey was performed on 4/15/2025, the laboratory was found not in compliance with the following CONDITION LEVEL DEFICIENCIES D2016 - 42 C.F.R. 493.803 Condition: Successful participation [proficiency testing]; D6000 - 42 C.F.R. 493.1403 Condition: Laboratories performing moderate complexity testing; laboratory director.
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on review of the Certification and Survey Provider Enhanced Reporting (CASPER) - 0155D and American Proficiency Institute (API) records (2023-2 and 2023-3) the laboratory failed to successfully participate in a proficiency testing</p>

	<p>program approved by HHS for each specialty, subspecialty and analyte or test in which the laboratory is certified under CLIA, the laboratory failed to successfully participate in the specialty of Hematology and analytes: Cell ID, RBC, HCT, HGB, WBC Count, Platelets resulting in unsuccessful performances. Refer to D2130 and D2131.</p>
<p>D2130</p>	<p>HEMATOLOGY CFR(s): 493.851(f)</p> <p>(f) Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on desk review of Certification and Survey Provider Enhanced Reporting (CASPER) Report 0155D Individual Laboratory Profile and API Evaluation report, the laboratory failed to achieve satisfactory performances for two of three consecutive proficiency events in 2024 for the following analytes: RBC - 2024 second testing event - 0%, 2024 third testing event - 0%; HCT - 2024 second testing event - 0%, 2024 third testing event - 0%; HGB - 2024 second testing event - 0%, 2024 third testing event - 0%; WBC Count - 2024 second testing event - 0%, 2024 third testing event - 0%; Platelets - 2024 second testing event - 0%, 2024 third testing event - 0%. A review of the 2024 proficiency testing scores from API confirmed the above findings.</p>
<p>D2131</p>	<p>HEMATOLOGY CFR(s): 493.851(g)</p> <p>(g) Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on review of the Certification and Survey Provider Enhanced Reporting (CASPER) Report 0155D Individual Laboratory Profile, CAP report, the laboratory failed to achieve satisfactory performance for two out of three consecutive events (2024-2 and 2024-3) for the specialty Hematology: Hematology - 2024 second testing event - 0%, 2024 third testing event - 0%. A review of the 2024 scores from api confirmed the above findings.</p>
<p>D6000</p>	<p>MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on a proficiency testing desk review of the Certification and Survey Provider</p>

Enhanced Reporting (CASPER) - 0155D report API records for 2024-2 and 2024-3 events, the laboratory director failed to provide overall management and a direction of the laboratory services. Refer to D6016.

D6016

LABORATORY DIRECTOR RESPONSIBILITIES

CFR(s): 493.1407(e)(4)(i)

(e)(4)(i) The proficiency testing samples are tested as required under Subpart H of this part;

This STANDARD is not met as evidenced by:
Based on a proficiency testing desk review of Certification and Survey Provider Enhanced Reporting (CASPER) - 0155D report API records for 2024-2 and 2024-3 events, the laboratory director failed to ensure successful participation in an HHS approved proficiency testing program. Refer to D2130 and D2131.