

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  05D0547058	<b>(X3) Date Survey Completed</b>  04/08/2021
<b>Name of Provider or Supplier</b>  Protzel Pathology Laboratory	<b>Street Address, City, State</b>  9735 Wilshire Blvd Ste 241, Beverly Hills, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on the surveyor's review of the laboratory's testing reagent expiration dates, and an interview with laboratory personnel (LP) on 4/8/2021 between 10 a.m and 11:30 a. m., it was determined that one reagent used for tissue staining was expired. Findings include: 1. On 4/8/2021, an inspection was conducted between 10 a.m. and 11:30 a.m. 2. During a review of the reagent expiration dates , it was noted at approximately 10: 30 a.m. that the expiration date on one container of the reagents used in the tissue staining process had expired. The LP recognized that this reagent had indeed expired. This reagent was currently in use for slide preparation and diagnostic evaluation. 3. The reagent that was expired was toluidine blue (used with the Cryostat) and it had expired on 2/1/2019.</p>