

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 05D0550065	(X3) Date Survey Completed 12/14/2022
Name of Provider or Supplier William A Pullen Md	Street Address, City, State 1301 20th St, Ste 550, Santa Monica, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2088	<p>ROUTINE CHEMISTRY CFR(s): 493.841(b)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on desk review of CMS PT records (CMS CASPER Report 0155D and 0153D, it was determined that the laboratory failed to achieve satisfactory performance for the same analyte or test in two out of three consecutive PT events for the analyte, Unsaturated Iron Binding Capacity (UIBC), resulting in an unsuccessful performances. The findings include: a. The laboratory failed to maintain successful performance with the PT program by failing to obtain a score of 80% of acceptable responses in two out of three consecutive PT events for the analyte, UIBC, as follows: 2022 Q1 2022 Q3 UIBC 40% 60% Q1 = First Testing Event Q3 = Third Testing Event b. Failure to achieve satisfactory performance for the same analyte or test for the analyts, UIBC.</p>