

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 05D0550324	(X3) Date Survey Completed 06/18/2019
Name of Provider or Supplier Westside Internal Medicine Laboratory	Street Address, City, State 2121 Wilshire Blvd, 3rd Fl Lab, Santa Monica, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on the lack of test procedural manual documentation and an interview with the Technical Consultant, the laboratory failed to have written instructions for proficiency testing samples to be followed by laboratory personnel. Findings include: a. The laboratory was unable to provide written instructions for proficiency testing samples. b. The Technical Consultant confirmed (May 1, 2019; 10:15 A.M.) that the laboratory did not have written instructions for the preanalytic, analytic, and postanalytic steps unique to proficiency testing program.</p>