Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 05D0569189	(X3) Date Survey Completed 07/13/2021
Name of Provider or Supplier Planned Parenthood	**Street Address, City, State** 2017 First Ave, Ste 100, San Diego, CA	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2153	ABO GROUP AND D(RHO) TYPING CFR(s): 493.859(a) Failure to attain a score of at least 100 percent of acceptable responses for each analyte or test in each testing event is unsatisfactory analyte performance for the testing event. This STANDARD is not met as evidenced by: Based on review on 07/13/2021 at 1:30 p.m. of the laboratory API (American Proficiency Institute) proficiency testing (PT) results reports, and interview with the laboratory staff, ten (10) random patient test reports reviewed from 12/32/2019 to 07/10/2021, it was determined that the laboratory failed to attain a score of at least 100 percent of acceptable responses for each analyte or test in each testing event was unsatisfactory analyte performance for the testing event. The findings included: a. The laboratory performed Rh (D) type and is enrolled in API (American Proficiency Institute) proficiency testing (PT) program using the Eldon card procedure. b. The laboratory attained a score of 60% for Rh (D) in the 1st PT event 2020 which was unsatisfactory analyte performance. Sample# RPT Result EXP Result RH-01 Rh Pos Rh Pos RH-02 Rh Pos Rh Neg RH-03 Reported problem Not Graded Rh-04 Rh Pos Rh Neg Rh-05 Reported problem Not graded c. The laboratory staff affirmed on 07/13/202 3:00 p.m. (survey date) that the laboratory failed to attain a score of at least 100 percent of acceptable responses for each analyte or test in the testing event. d. The laboratory's testing declaration estimates the annual volume of Rh typing performed and reported at 3,000 tests.