

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  05D0596124	<b>(X3) Date Survey Completed</b>  01/06/2026
<b>Name of Provider or Supplier</b>  North East Medical Services Inc	<b>Street Address, City, State</b>  2171 Junipero Serra Blvd, Ste 400, Daly City, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2087</b>	<p><b>ROUTINE CHEMISTRY</b> CFR(s): 493.841(a)</p> <p>(a) Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of the AAB-MEDICAL LABORATORY EVALUATION proficiency testing records, and interview with the Technical Supervisor (TS) on January 6, 2026, it was determined that the laboratory failed to attain a score of at least 80 percent of acceptable responses for routine chemistry for each analyte in 2024 and 2025. The findings included: 1. The laboratory enrolled in the AAB-MEDICAL LABORATORY EVALUATION proficiency testing program for routine chemistry testing using Abbott Alinity analyzer. According to the AAB evaluation, the laboratory attained the following unsatisfactory scores: Q1- 2024: CREATININE = 40% Q1- 2025: CREATININE = 60% Q3- 2024: ALBUMIN = 40% 2. On January 6, 2026, at approximately 10:00 am, the TS confirmed that the laboratory received the above unsatisfactory proficiency scores. 3. The laboratory's testing declaration form, signed by the laboratory director on December 10, 2025, stated that the laboratory performed approximately 86,804 routine chemistry tests including albumin and creatinine annually.</p>