

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  05D0600646	<b>(X3) Date Survey Completed</b>  06/23/2021
<b>Name of Provider or Supplier</b>  Sierra Hematology And Oncology	<b>Street Address, City, State</b>  1580 Creekside Dr Ste 230, Folsom, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2121</b>	<p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of the third quarter (Q3-2020) of the American Proficiency Institute (API) performance summary report, ten (10) randomly selected random patient sampling records from 12/03/2019 to 05/28/2021 and interview with the laboratory personnel on 06/23/2021 at 9:30 a.m., it was determined that the laboratory failed to attain a score of at least 80 percent of acceptable responses for Red Cell Counts (RBC). The findings included: 1. Q3-2020, API reported the following unsatisfactory proficiency testing scores. Analyte: Score: Event/Year: Performance: RBC 0% Q3 /2020 Unacceptable Sample Reported Results Expected Results HEM-11 3.75 3.80-4.30 HEM-12 4.72 4.74-5.36 HEM-13 2.00 2.03-2.30 HEM-14 4.67 4.69-5.30 HEM-15 3.36 3.38-3.82 2. The laboratory resulted and reported patient Red Cell Counts (RBC) tests during the approximate time the laboratory received the API (Q3-2020) unsatisfactory scores. 3. The testing personnel affirmed on 06/23/2021, 11:00 a.m. (survey date) that the laboratory received the above unsatisfactory scores. 4. Based on the laboratory's annual testing declaration 06/21/2021 approximately 2,808 RBC patient tests were resulted and reported</p>