

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 05D0615137	(X3) Date Survey Completed 07/21/2025
Name of Provider or Supplier Advanced Cosmetic Surgery & Dermatology	Street Address, City, State 276 Kingsbury Grade Ste 101, Stateline, NV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	This Statement of Deficiencies was created as a result of an on-site CLIA recertification survey conducted at your facility on July 21, 2025. The findings and conclusions of any investigation by the Division of Public and Behavioral Health shall not be construed as prohibiting any criminal or civil investigations, actions or other claims for relief that may be available to any party under applicable federal, state, or local laws.
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the 2024 laboratory records for the twice per year verification of accuracy for KOH testing, and an interview with the office manager, the laboratory failed to ensure that the twice per year verification of accuracy was performed for each provider performing KOH testing in the laboratory. Findings include: 1. Testing personnel number one on the CMS 209 form did not have documentation of accuracy checks in 2024. 2. An interview with the office manager on July 21, 2025 at approximately 12:00 PM confirmed these findings. The laboratory performs approximately 7 mycology and parasitology tests annually.</p>
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p>

This STANDARD is not met as evidenced by:

Based on observation of the testing supplies available during a walk-through of the facility on July 21, 2025, at approximately 12:30 PM, and an interview with the office manager, the laboratory failed to ensure that testing supplies available for use were not expired. Findings include: 1. The laboratory had expired KOH testing supplies available for use. The KOH bottle (lot number 1120) expired on April 30, 2023. There were no KOH bottles available that was not expired. 2. An interview with the office manager confirmed these findings. This is a repeat deficiency cited during the recertification survey performed on December 15, 2023. The laboratory performs approximately 7 mycology and parasitology tests annually.