

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 05D0617090	<b>(X3) Date Survey Completed</b> 03/10/2022
<b>Name of Provider or Supplier</b> Butte County Health Dept Laboratory	<b>Street Address, City, State</b> 695 Oleander Ave, Chico, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2047</b>	<p>PARASITOLOGY CFR(s): 493.829(a)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: The STANDARD is not met as evidenced by: Based on the surveyor's review of the laboratory's records for evaluation of proficiency testing performance and an interview with laboratory personnel (LP) on 3/10/22 between 1:00 p.m. and 3:00 p.m, it was determined that the Parasitology proficiency test results were below 80 % for cycle 2 in 2020. Findings include: 1. On 3/10/22, an inspection was conducted between 1:00 p.m. and 3:00 p.m. 2. During a review of the laboratory documentation from API (American Proficiency Institute- the agency providing the proficiency specimens), it was noted at approximately 2:00 p.m. that the laboratory had unacceptable Parasitology results for cycle 2 in 2020. 3. The findings and acceptable values were as follows: Test: Giardia/Cryptosporidium Antigens Sample Actual Result Expected Result Crypto. Ag CRG-06 Positive Negative Giardia Ag CRG-07 Negative Positive Test Group: Parasitology Sample Actual Result Expected Result PAR-08 Iodamoeba but. Giardia duod. PAR-9 Giardia duod. Iodamoeba but. 4. The LP recognized the above results.</p>