

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 05D0644240	(X3) Date Survey Completed 03/09/2022
Name of Provider or Supplier Shasta County Public Health Lab	Street Address, City, State 2650 Breslauer Way, Redding, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5775	<p>COMPARISON OF TEST RESULTS CFR(s): 493.1281(a)(c)</p> <p>(a) If a laboratory performs the same test using different methodologies or instruments, or performs the same test at multiple testing sites, the laboratory must have a system that twice a year evaluates and defines the relationship between test results using the different methodologies, instruments, or testing sites. (c) The laboratory must document all test result comparison activities.</p> <p>This STANDARD is not met as evidenced by: Based on the surveyor's review of the laboratory's records for parallel instrument testing and an interview with the laboratory director (LD) on 3/9/22 between 9:00 a.m. and 12:00 p.m, it was determined that there was not documentation for parallel testing in 2020 for two instruments. Findings include: 1. On 3/9/22, an inspection was conducted between 9:00 a.m. and 12:00 p.m. 2. During a review of the laboratory documentation to support parallel testing (parallel testing is required each year when two instruments are used to report results for the same analyte), it was noted that there was not parallel testing data for the two Applied Biosystem 7500 (ABI) devices in 2020. 3. The LD recognized the above results.</p>