Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 05D0699248	**(X3) Date Survey Completed** 07/29/2021
Name of Provider or Supplier C-Ton Laboratory	**Street Address, City, State** 3870 Del Amo Blvd, Ste 504, Torrance, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D2087	ROUTINE CHEMISTRY CFR(s): 493.841(a) Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event. This STANDARD is not met as evidenced by: The STANDARD is not met as evidenced by: Based on the surveyor's review of the laboratory's records for evaluation of proficiency testing performance from CMS, laboratory proficiency reporting from AAB (American Association of Bioanalysts) and an interview with the laboratory Technical Supervisor (TS) on 7/29/2021 between 10:00 a.m. and 12:30 p.m., it was determined that there were two unacceptable (below 80 %) proficiency testing (PT) results for Prostate Specific Antigen (PSA) for Cycles 1 and 2 in 2021. Findings include: 1. On 7/29/2021, an inspection was conducted between 10:00 a.m. and 12:30 p.m. 2. During a review of the laboratory documentation from AAB (agency providing the proficiency specimens), it was noted at approximately 11 a.m. that there were two unacceptable PT results for Cycles 1 and 2 for PSA on the 2021 proficiency reports. The laboratory utilizes the DXI instrument for chemistry testing. The TS recognized these atypical results. 3. The findings and acceptable ranges were as follows: Analyte: PSA Cycle 1 2021 Sample Actual Result Expected Result (range) 2 22.7 23.1-34.7 Cycle 2 2021 Sample Actual Result Expected Result (range) 1 20.9 22.1-33.2