

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 05D0867202	(X3) Date Survey Completed 03/03/2026
Name of Provider or Supplier Los Angeles Hematology Oncology	Street Address, City, State 1801 W Romneya Drive Ste 203, Anaheim, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2121	<p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>(a) Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of the AAB-MEDICAL LABORATORY EVALUATION proficiency testing (PT) records, and interview with the laboratory technical consultant (TC) on March 03, 2026, it was determined that the laboratory failed to attain a score of at least of 80 percent of acceptable responses for the analyte Red Blood Cell (RBC) count for the first PT event of 2025. The findings included: 1. The laboratory enrolled in the AAB-MEDICAL LABORATORY EVALUATION proficiency testing program for Hematology testing using Abbott Cell-Dye 1800 analyzer. According to the PT evaluation report, the laboratory attained an unsatisfactory score of 40% for RBC count in first PT event of 2025. 2. On March 3, 2026, at approximately 11:30 am, the laboratory technical consultant affirmed that the laboratory received the above unsatisfactory proficiency scores. 3. The laboratory's testing declaration form, signed by the laboratory director on March 2, 2026, stated that the laboratory performed approximately 10,400 hematology tests including RBC annually.</p>
D2123	<p>HEMATOLOGY CFR(s): 493.851(c)</p> <p>(c) Failure to participate in a testing event is unsatisfactory performance and results in a score of 0 for the testing event. Consideration may be given to those laboratories failing to participate in a testing event only if-- (1) Patient testing was suspended during the time frame allotted for testing and reporting proficiency testing results; (2)</p>

The laboratory notifies the inspecting agency and the proficiency testing program within the time frame for submitting proficiency testing results of the suspension of patient testing and the circumstances associated with failure to perform tests on proficiency testing samples; and (3) The laboratory participated in the previous two proficiency testing events.

This STANDARD is not met as evidenced by:

Based on review of the AAB-MEDICAL LABORATORY EVALUATION proficiency testing (PT) records, and an interview with the laboratory technical consultant (TC) on March 03, 2026, it was determined that the laboratory failed to participate in the third testing event of 2024 for the specialty of Hematology. The findings included: 1. The laboratory enrolled in the AAB-MEDICAL LABORATORY EVALUATION proficiency testing program for Hematology testing using Abbott Cell-Dye 1800 analyzer. According to the PT evaluation report, the laboratory attained score of 0 % in Hematology in the third testing event of 2024. 2. On March 3, 2026, at approximately 11:30 am, the laboratory technical consultant affirmed that the laboratory received the above unsatisfactory proficiency scores. 3. The laboratory's testing declaration form, signed by the laboratory director on March 2, 2026, stated that the laboratory performed approximately 10,400 hematology tests annually.