

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 05D0936846	(X3) Date Survey Completed 03/25/2026
Name of Provider or Supplier Pacific Reproductive Center	Street Address, City, State 3720 Lomita Blvd Ste 200, Torrance, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1) At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. This STANDARD is not met as evidenced by: Based on observation of the Beckman Coulter Access II analyzer (serial number 572646), review of laboratory records for 2024-2025, the lack of records, and interview with the Testing Person/Technical Consultant, it was determined the laboratory failed to verify the accuracy of testing for FSH (Follicle Stimulating Hormone), Prolactin, TSH (Thyroid Stimulating Hormone), and AMH (Anti-Muellerian Hormone). Findings included: 1. The laboratory chose to enroll in AAB /MLE's proficiency testing program as the means to verify the accuracy of testing for FSH and Prolactin. 2. Laboratory proficiency test records for 2024 failed to verify test accuracy, as follows: a. FSH i. Events 1 and 2: No scores/ No participation. ii. For event 3, the laboratory reported 1 unacceptable result out of 2 for a score of 50% b. Prolactin i. Event 2: Lack of consensus; reported results could not be evaluated for accuracy. ii. Event 3: the laboratory reported 1 unacceptable result out of 2 for a score of 50%. 3. The laboratory didn't participate in AAB/MLE proficiency testing for TSH or AMH in 2024 and 2025. 4. The laboratory had no other records for verifying the accuracy of testing for FSH, Prolactin, TSH, or AMH. 5. The Testing Person /Technical Consultant affirmed (3/25/26 at 1:00 PM) the aforementioned findings and that no alternate methods were utilized to verify the accuracy of testing. 6. The accuracy of testing was not verified for results reported during 2024, as follows: i. FSH: 19 out of 19 results reported during the timeframe 2/16/24 - 12/13/24; ii. Prolactin: 13 out of 20 results reported 2/16/24 - 12/04/24; 7. The accuracy of testing was not verified for results reported during 2024 -2025, as follows i. TSH: 8 out of 8 results reported 3/18/24 - 10/07/24; 1 out of 1 reported 7/09/25. ii. AMH: 22 out of 22 results reported 3/18/24 - 3/21/25. .