

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 05D0941947	(X3) Date Survey Completed 04/25/2019
Name of Provider or Supplier Central Coast Multispecialty	Street Address, City, State 9833 Blue Larkspur Ln, Monterey, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2098	<p>ENDOCRINOLOGY CFR(s): 493.843(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of 2017: Event 3 proficiency testing reports from CMS (Report 155D, Individual Laboratory Profile) and AAB (American Association of Bioanalysts), laboratory proficiency testing records, and patients test records; and interview with the Laboratory Director (Testing person), it was revealed that the laboratory failed to attain a score of at least 80% for TSH (Thyroid Stimulating Hormone). Findings included: a. CMS and API reported a score of 0% based on the laboratory's 5 unacceptable results out of 5. b. The Laboratory Director (Testing person) affirmed (4/25/19 at 5:30pm) the aforementioned score; and thus unsatisfactory testing for TSH. c. The reliability and quality of results reported for TSH could not be assured when testing was unsatisfactory. Based on review of patients tests records, the laboratory reported no TSH results in September 2017 and fewer than 2 during the timeframe October - December 2017.</p>