

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  05D0941947	<b>(X3) Date Survey Completed</b>  10/06/2021
<b>Name of Provider or Supplier</b>  Central Coast Multispecialty	<b>Street Address, City, State</b>  9833 Blue Larkspur Ln, Monterey, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2098</b>	<p><b>ENDOCRINOLOGY</b> CFR(s): 493.843(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on reviews of analyte performance of the American Association of Bioanalysts (AAB) proficiency testing performance summary records, six (6) randomly selected patients' test requisitions from 02/27/2020 to 10/04/2021 and test results (medical records), quality control reviewed, proficiency testing (PT) result reports and interview with the laboratory director on 10/06/2121 at 10:30 a.m., it was determined that the laboratory failed to attain a score of at least 80 percent of acceptable responses for 3rd quarter (Q3) proficiency event of 2020 for thyroid stimulating hormone (TSH). The findings included: 1. AAB Q3 2020 reported Proficiency Testing Event score of 0% for TSH: Event Analyte Sample Results Exp. Results Q3 2020 TSH 11 4.5 5.7-7.6 Q3 2020 TSH 12 1.7 2.1-2.9 Q3 2020 TSH 13 1.1 1.3-1.9 Q3 2020 TSH 14 5.1 5.7-8 Q3 2020 TSH 15 3.6 4.1-5.9 2. Patient test results reviewed covering the period of Q3, 2020, the laboratory analyzed and reported TSH tests during this time. 3. The laboratory director confirmed 10/06/2021, 11:30 a.m. (survey date) that the laboratory received the above unsatisfactory proficiency testing score. 4. The laboratory annual testing declaration (09/30/2021) for Endocrinology (TSH included) estimated that a total test volume of 400.</p>