

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  05D0941947	<b>(X3) Date Survey Completed</b>  11/21/2024
<b>Name of Provider or Supplier</b>  Central Coast Multispecialty	<b>Street Address, City, State</b>  9833 Blue Larkspur Ln, Monterey, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2098</b>	<p>ENDOCRINOLOGY CFR(s): 493.843(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on interview with Laboratory Director (LD) and review of Proficiency Testing (PT) evaluation records on November 21, 2024, it was determined that the Laboratory failed to attain a score of at least 80% of acceptable responses for Thyroid-stimulating Hormone (TSH) in the third events of 2022 and 2023, which were unsatisfactory analyte performance for the testing events. The findings include: 1. The laboratory enrolled in the American Association of Bioanalysts (AAB) proficiency testing program for TSH testing using Tosoh AIA 900 instrument in 2022, 2023 and 2024. According to the AAB evaluation, the laboratory attained a score of 40% in the third event of 2022 and 60 % in the third event of 2023 for TSH: Event Analyte Sample Repor. Value Exp. Results Q3 2022 TSH 11 6.7 4.1 - 6.7 Q3 2022 TSH 12 3.1 1.9 - 2.8 Q3 2022 TSH 13 5.7 3.5 - 5.4 Q3 2022 TSH 14 9.9 4.2 -11.5 Q3 2022 TSH 15 7.9 4.7-7.8 Q3 2023 TSH 11 9.6 6.1 - 9.7 Q3 2023 TSH 12 3.0 1.9 - 3.1 Q3 2023 TSH 13 8.1 5.4 - 7.5 Q3 2023 TSH 14 6.8 4.5 -6.6 Q3 2023 TSH 15 5.7 3.5 -5.8</p> <p>2. On November 21, 2024, at approximately 10:00 am the laboratory director affirmed that the laboratory received the above unsatisfactory proficiency scores. 3. The laboratory's testing declaration form, signed by the laboratory director on October 28, 2024, stated that the laboratory performed approximately 280 endocrinology tests annually.</p>