

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 05D0956168	(X3) Date Survey Completed 04/30/2026
Name of Provider or Supplier 21st Century Diagnostic Laboratory	Street Address, City, State 121 N Victory Blvd, Burbank, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A proficiency testing desk review survey as performed on April 30, 2026, the laboratory was found not in compliance with the following CONDITION LEVEL DEFICIENCIES D2016 - 42 C.F.R. 493.803 Condition: Successful [proficiency testing] participation; and D6076 - 42 C.F.R. 493.1441 Condition: Laboratories performing high complexity testing; laboratory director.
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on review of the Certification and Survey Provider Enhanced Reporting (CASPER) - 0155D and the American Association of Bioanalysts - Medical Laboratory Evaluation (AAB-MLE) records for 2023-2 and 2023-3 & 2025-2 and

	2026-1 events, the laboratory failed to successfully participate in a proficiency testing program approved by HHS for each specialty, subspecialty and analyte or test in which the laboratory is certified under CLIA, the laboratory failed to successfully participate in Bacteriology resulting in unsuccessful performances. See D2028.
D2028	BACTERIOLOGY CFR(s): 493.823(e) (e) Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on review of the Certification and Survey Provider Enhanced Reporting (CASPER) Report 0155D Individual Laboratory Profile and American Association of Bioanalysts - Medical Laboratory Evaluation (AAB-MLE) report, the laboratory failed to achieve an overall satisfactory proficiency performances for 2023-2 and 2023-3 & 2025-2 and 2026-1 for Bacteriology (specialty Microbiology) resulting into subsequent failure: The findings include: 1. Bacteriology 48% - 2023 second testing event; Bacteriology 43% - 2023 third testing event; Bacteriology 55% - 2025 second testing event; Bacteriology 40% - 2026 first testing event. A review of the 2023, 2025 and 2026 scores from the American Association of Bioanalysts - Medical Laboratory Evaluation (AAB-MLE) confirmed the above findings.
D6076	LABORATORY DIRECTOR CFR(s): 493.1441 The laboratory must have a director who meets the qualification requirements of 493.1443 of this subpart and provides overall management and direction in accordance with 493.1445 of this subpart. This CONDITION is not met as evidenced by: Based on a proficiency testing desk review of the CASPER 0155D report and American Association of Bioanalysts - Medical Laboratory Evaluation (AAB-MLE) records for 2023-2 and 2023-3 & 2025-2 and 2026-1 Bacteriology events, the laboratory director failed to provide overall management and direction of the laboratory services. Refer to D6089.
D6089	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(4)(i) (e)(4)(i) The proficiency testing samples are tested as required under subpart H of this part; This STANDARD is not met as evidenced by: Based on a proficiency testing desk review of the CASPER 0155D report and American Association of Bioanalysts - Medical Laboratory Evaluation (AAB-MLE) records for 2023-2 and 2023-3 & 2025-2 and 2026-1 Bacteriology events, the laboratory director failed to ensure successful participation in an HHS proficiency testing program. Refer to D2028 analyte deficiency/deficiencies.