

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  05D0957076	<b>(X3) Date Survey Completed</b>  12/12/2018
<b>Name of Provider or Supplier</b>  Diagnostic Laboratory Science Inc	<b>Street Address, City, State</b>  7711 Garden Grove Blvd, Garden Grove, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2087</b>	<p><b>ROUTINE CHEMISTRY</b> CFR(s): 493.841(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of the second quarter (Q2-2018) of the American Association of Bioanalysts (AAB) proficiency testing records, random patient sampling test results, and interview with the technical consultant, it was determined that the laboratory failed to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event. The findings included: a. Q2-2018, AAB reported 60% Total Thyroxine (TT4) analyte. b. For five (5) out of five (5) random patient sampling test results reviewed covering period from 12/30/2016 to 10/10/2018, the laboratory analyzed and reported TT4 analyte during the time the laboratory received an unsatisfactory proficiency testing score. c. The technical consultant confirmed (12/12/2018, 1600), that the laboratory received the above unsatisfactory proficiency test score.</p>
<b>D2098</b>	<p><b>ENDOCRINOLOGY</b> CFR(s): 493.843(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of the second quarter (Q2-2016) of the American Association of Bioanalysts (AAB) proficiency testing records, random patient sampling test results, and interview with the technical consultant, it was determined that the laboratory</p>

failed to a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event. The findings included: a. Q2-2016, API reported 60% for Blood Urea Nitrogen (BUN) analyte. b. For one (1) out of one (1) random patient sampling test results reviewed 11/26/2016, the laboratory analyzed and reported BUN analyte during the time the laboratory received an unsatisfactory proficiency testing score. c. The technical consultant confirmed (12/12/2018, 1600), that the laboratory received the above unsatisfactory proficiency test score.