

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  05D0970561	<b>(X3) Date Survey Completed</b>  10/11/2018
<b>Name of Provider or Supplier</b>  Metrolab, Inc	<b>Street Address, City, State</b>  16550 Ventura Blvd Ste 402, Encino, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5217</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of the third quarter (Q3-2016), first quarter (Q1-2017), second quarter (Q2-2017), and second quarter (Q2-2018) of the American Association of Bioanalysts (AAB), proficiency testing records, and interview with the testing personnel, it was determined that the laboratory failed to attain at least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. The findings included: a. Q3-2016, Q1-2017, Q2-2017, and Q2-2018, AAB reported the following unsatisfactory proficiency testing scores. Analyte: Score: Event/Year: Folate 50% Q3-2016 Testos 50% Q1-2017 PTH 50% Q2-2017 CA-125 50% Q2-2018 Testos 0% Q2-2018 Folate 50% Q2-2018 b. Based on the laboratory's testing declaration submitted for 2017 to 2018, the laboratory analyzed and reported 1,200 Folate, 1,316 Testosterone, 379 PTH, and CA-125 patient test results that cannot be assured. d. The testing personnel confirmed (10/11/2018, 1800) that the laboratory received the above unsatisfactory proficiency testing scores.</p>