Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 05D1077766	**(X3) Date Survey Completed** 11/22/2021
Name of Provider or Supplier Beverly Hills Cancer Center	**Street Address, City, State** 8900 Wilshire Blvd, Beverly Hills, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D2087	ROUTINE CHEMISTRY CFR(s): 493.841(a) Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event. This STANDARD is not met as evidenced by: Based on the surveyor's review of the laboratory's records for evaluation of proficiency testing performance from the CMS 96 Report, laboratory proficiency reporting from API, and an interview with the laboratory technical consultant (TC) on 11/22/2021 between 9:30 a.m. and 11:30 a.m., it was determined that there were three unacceptable (below 80 %) proficiency testing (PT) results for Chemistry in the 2020 and 2021 submission date ranges. Findings include: 1. On 11/22/2021, an inspection was conducted between 9:30 a.m. and 11:30 a.m. 2. During a review of the laboratory documentation from API (agency providing the proficiency specimens), it was noted at approximately 10 a.m. that there were three unacceptable PT results for general chemistry. The laboratory utilizes the OCD Vitros instruments for general chemistry testing. The TC recognized these atypical results. 3. The findings and acceptable ranges were as follows: Cycle 2 2020 Chemistry Analyte: TIBC (3/5 unacceptable) Sample Actual Result Expected Result (range) CH-06 317 327-502 CH-09 206 209-292 CH-10 285 302-392 Cycle 2 2021 Chemistry Analyte: Glucose (5/5 unacceptable) Sample Actual Result Expected Result (range) CH-06 84 63-78 CH-07 131 100-123 CH-08 202 156-192 CH-09 177 137-168 CH-10 236 186-229 Cycle 3 2020 Chemistry Analyte: Ferritin (2/2 unacceptable) Sample Actual Result Expected Result (range) IA-11 78 46-59 IA-12 122 69-100 4. The TC affirmed the unacceptable results listed above.
D6022	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(5)

	The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(5) Ensure that the quality control and quality assessment programs are established and maintained to identify failures in quality as they occur. This STANDARD is not met as evidenced by: Based on the surveyor's review of the laboratory's records for evaluation of proficiency testing performance from the CMS 96 Report, laboratory proficiency reporting from API, and an interview with the laboratory technical consultant (TC) on 11/22/2021 between 9:30 a.m. and 11:30 a.m., it was determined that there were three unacceptable (below 80 %) proficiency testing (PT) results for Chemistry in the 2020 and 2021 submission date ranges. . The findings include: See D-2087 (3 deficiencies) listed above.