

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 05D1089660	(X3) Date Survey Completed 10/12/2018
Name of Provider or Supplier Universal Pain Management	Street Address, City, State 819 Auto Center Dr, Palmdale, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of first quarter (Q1-2016), second quarter (Q2-2016), second quarter (Q2-2017) of the American Proficiency Institute (API) proficiency testing records, laboratory's proficiency testing results, random patient sampling test results, and interview with the laboratory director, it was determined that the laboratory failed to at least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. The findings included: a. API reported the following unsatisfactory proficiency testing scores. Analyte: Score: Event/Year: UDS Alcohol 50% Q1-2016 UDS Barbiturates 50% Q1-2016 UDS Benzo- diazepines 50% Q1-2016 UDS Canna- binoids 50% Q1-2016 UDS Methadone 50% Q1-2016 UDS Oxycodone 50% Q1-2016 UDS Methadone 50% Q2-2016 UDS Methadone 33% Q2-2017 b. For twelve (12) out of twelve (12) random patient test results reviewed covering period from 1/8/2016 to 4/28/2018, the laboratory analyzed and reported during the periods where in the laboratory received an unsatisfactory proficiency testing scores. c. The laboratory director confirmed (10 /12/2018, 1300) that the laboratory received the above unsatisfactory proficiency testing scores for the above analytes.</p>