

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 05D2060295	(X3) Date Survey Completed 09/10/2020
Name of Provider or Supplier San Diego Family Dermatology	Street Address, City, State 15725 Pomerado Rd Ste 102, Poway, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on the surveyor's review of the laboratory's records and an examination of laboratory reagents, along with an interview with laboratory personnel (LP) on 9/10/2020 between 2 pm and 3:30 pm, it was determined that the laboratory failed to not use a reagent when the expiration date had passed. Findings include: 1. On 9/10/2020, an inspection was conducted between 2 pm and 3:30 pm. 2. During a review of the laboratory documentation and an examination of the KOH reagent, it was noted at approximately 3 pm that the expiration date on the reagent bottle had passed. The LP recognized that the reagent date had passed. 3. The reagent was KOH from McKesson, and the expiration date on the bottle was 10/2/2018.</p>