

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 05D2077707	(X3) Date Survey Completed 12/18/2018
Name of Provider or Supplier Uc San Diego Health -	Street Address, City, State 910 Sycamore Ave Ste 102, Vista, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2121	<p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of first quarter (Q1-2017) of the College of American Pathologists (CAP) proficiency testing records, CMS Casper 0096D report, random patient sampling test results, and interview with the technical consultant, it was determined that the laboratory failed to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event. The findings included: a. Q1-2017 CAP reported a 0% score for Hematology proficiency testing. b. For seven (7) out of seven (7) random patient sampling test results reviewed covering period from 9/9 /2016 to 9/27/2018, the laboratory analyzed and reported Complete Blood Count (CBC) tests during the approximate period where in the laboratory received the unsatisfactory proficiency testing score of 0%. c. The technical consultant confirmed (12/18/2018, 1430), that the laboratory received the above unsatisfactory proficiency testing score.</p>