

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 05D2093483	(X3) Date Survey Completed 05/16/2024
Name of Provider or Supplier Elite Urgent Care Inc	Street Address, City, State 1524 Mchenry Ave, Ste 130, Modesto, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A proficiency testing desk review survey was performed on May 16, 2024, the laboratory was found not in compliance with the following CONDITION LEVEL DEFICIENCIES D2016 - C.F.R. 493.803 Conditions: Successful participation [proficiency testing] D6000 - 42 C.F.R. 493.1403 Condition: Laboratories performing moderate complexity testing; Laboratory Director.
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on review of the Certification and Survey Provider Enhanced Reporting (CASPER) -0155D and WSLH Proficiency Testing records (2023-1, 2023-3 and 2024-1), the laboratory failed to successfully participate in a proficiency testing program

	approved by HHS, for each specialty, subspecialty or test which the laboratory is certified under CLIA. The laboratory failed to successfully participate in the specialty Hematology for the analyte - white cell differential (WBC DIFF) . Refer to D2130.
D2130	HEMATOLOGY CFR(s): 493.851(f) Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on review of the Certification and Survey Provider Enhanced Reporting (CASPER) Report 0155D Individual Laboratory Profile and WSLH Proficiency Testing report, the laboratory failed to achieve satisfactory performance in three of four events events (2023-1, 2023-3, and 2024-1) for the analyte WBC DIFF. The finding include: a. 4% on the 2023 first event 16% on the 2023 third event 16% on the 2024 first event b. A review of the 2023 and 2024 scores from WSLH Proficiency Testing confirmed the above findings.
D6000	MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403 The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart. This CONDITION is not met as evidenced by: Based on a proficiency testing desk review of the CASPER 0155D report and WSLH Proficiency Testing records for 2023-1, 2023-3, and 2024-1 events, the laboratory director failed to provide overall management and direction of the laboratory services. Refer to D6016.
D6016	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i) The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part; This STANDARD is not met as evidenced by: Based on a proficiency testing desk review of CASPER 0155D report and WSLH Proficiency Testing records for 2023-1, 2023-3, and 2024-1 events, the laboratory director failed to ensure successful participation in an HHS approved proficiency testing program. Refer to D2130.