

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 05D2197767	(X3) Date Survey Completed 08/09/2022
Name of Provider or Supplier Sollis Health La, Pc	Street Address, City, State 155 N San Vincente Blvd 1st Fl, Beverly Hills, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2020	<p>BACTERIOLOGY CFR(s): 493.823(a)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on the surveyor's review of the laboratory's records for evaluation of proficiency testing performance and an interview with laboratory personnel (LP) on 8/9/22 between 2:00 p.m. and 3:30 p.m, it was determined that there were proficiency testing (PT) scores below 80 % for one testing cycle. The issue relates to Bordetella Parapertussis testing. Findings include: 1. On 8/9/22, an inspection was conducted between 2:00 p.m. and 3:30 p.m. 2. During a review of the laboratory documentation from AP (American Proficiency Institute- the agency providing the proficiency specimens), it was noted at approximately 2:30 p.m. that the laboratory had unacceptable Microbiology results as indicated below. The testing was performed on the BioFire FilmArray. 3. The findings and acceptable ranges were as follows: Cycle 1 2022 Test: Respiratory Bordetella Parapertusis Sample Actual Result Expected Result (range) Rsp-01 Detected Not Detected Rsp-02 Detected Not Detected Rsp-03 Detected Not Detected Rsp-04 Detected Not Detected Rsp-05 Detected Not Detected 4. The laboratory reporting testing for approximately 200 Respiratory Panels for a one year period, and Bordetella Pertusis is a component of the Biofire Respiratory panel. 5. The LP recognized the above results.</p>
D2075	<p>GENERAL IMMUNOLOGY CFR(s): 493.837(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p>

This STANDARD is not met as evidenced by:

Based on the surveyor's review of the laboratory's records for evaluation of proficiency testing performance and an interview with laboratory personnel (LP) on 8/9/22 between 2:00 p.m. and 3:30 p.m, it was determined that there were proficiency testing (PT) scores below 80 % for one testing cycle. The issue relates to C-Reactive Protein (CRP). Findings include: 1. On 8/9/22, an inspection was conducted between 2:00 p.m. and 3:30 p.m. 2. During a review of the laboratory documentation from AP (American Proficiency Institute- the agency providing the proficiency specimens), it was noted at approximately 3:00 p.m. that the laboratory had unacceptable Immunology results as indicated below. The testing was performed on the Abaxis Picollo instrument. 3. The findings and acceptable ranges were as follows: Cycle 1 2022 Test: C-Reactive Protein (CRP) Sample Actual Result Expected Result (range) Crp-01 22.4 2.15-2.97 Crp-02 5.00 0.00-0.30 4. The laboratory reporting testing for approximately 100 general Immunology tests to include CRP. 5. The LP recognized the above results.