

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 05D2274138	(X3) Date Survey Completed 08/14/2025
Name of Provider or Supplier Lifequest Diagnostic, Inc	Street Address, City, State 720 N Valley St Ste F, Anaheim, CA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2087	<p>ROUTINE CHEMISTRY CFR(s): 493.841(a)</p> <p>(a) Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's proficiency testing (PT) records, and interview with the Technical Consultant (TC) on August 14, 2025, it was determined that the laboratory failed to attain a score of at least 80 percent of acceptable responses for the Routine chemistry analyte Sodium for the third PT event of 2024. The findings included: 1. The laboratory participated in the AAB-MEDICAL LABORATORY EVALUATION (AAB-MLE) proficiency testing (PT) program for endocrinology using the Horiba Pentra C400 analyzer. According to the AAB-MLE evaluation report, the laboratory received an unsatisfactory score of 0% for Sodium in M3-2024. 2. On August 14, 2025, at approximately 1:30 pm the Technical Consultant affirmed that the laboratory received the above unsatisfactory proficiency scores. 3. The laboratory's testing declaration form, signed by the laboratory director on August 14, 2025, stated that the laboratory performed approximately 500 Routine chemistry tests including Sodium annually. Thus, the accuracy and reliability of patient test reports cannot be determined.</p>
D2098	<p>ENDOCRINOLOGY CFR(s): 493.843(a)</p> <p>(a) Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p>

This STANDARD is not met as evidenced by:

Based on review of the laboratory's proficiency testing (PT) records, and interview with the Technical Consultant (TC) on August 14, 2025, it was determined that the laboratory failed to attain a score of at least 80 percent of acceptable responses for the Endocrinology analyte Free Thyroxine for the first PT event of 2025. The findings included: 1. The laboratory participated in the AAB-MEDICAL LABORATORY EVALUATION (AAB-MLE) proficiency testing (PT) program for endocrinology using the TOSOH AIA-360 analyzer. According to the AAB-MLE evaluation report, the laboratory received an unsatisfactory score of 0% for Free Thyroxine in M1-2025. 2. On August 14, 2025, at approximately 1:30 pm the Technical Consultant affirmed that the laboratory received the above unsatisfactory proficiency scores. 3. The laboratory's testing declaration form, signed by the laboratory director on August 14, 2025, stated that the laboratory performed approximately 300 endocrinology tests including Free Thyroxine annually. Thus, the accuracy and reliability of patient test reports cannot be determined.