

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 06D0055977	(X3) Date Survey Completed 11/02/2021
Name of Provider or Supplier Haxtun Hospital District	Street Address, City, State 235 W Fletcher St, Haxtun, CO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on routine desk audit of CMS-153 and 155 reports of proficiency testing performance and interview, the laboratory failed to achieve satisfactory performance for pH Blood Gas for two consecutive events (2021 event 2 and 3), resulting in unsuccessful proficiency testing performance. See D2096."</p>
D2096	<p>ROUTINE CHEMISTRY CFR(s): 493.841(f)</p> <p>Failure to achieve satisfactory performance for the same analyte or test in two</p>

consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on review of proficiency testing scores and interview, the laboratory failed to achieve a score of 80 percent for pH of Blood Gas in the specialty of routine chemistry in two consecutive events in 2021, resulting in unsuccessful performance. The findings include: 1. During a review on October 26, 2021 of the CMS-153 Unsuccessful Proficiency Testing Report included Haxtun Hospital District with unsatisfactory proficiency testing scores for pH of Blood Gas. 2. During a review on October 26, 2021 of the CMS-155 report, the American Proficiency Institute (API) pH of Blood Gas score for event 2 of 2021 was 60%. 3. During a review on October 26, 2021 of the CMS-155 report, the American Proficiency Institute (API) pH of Blood Gas score for event 3 of 2021 was 60%. 4. On October 27, 2021, the laboratory manager stated that the laboratory was aware of the unsatisfactory proficiency test results and had performed the investigation and corrective action for this analyte.