

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 06D0512777	(X3) Date Survey Completed 06/15/2021
Name of Provider or Supplier Advanced Dermatology Of Colorado	Street Address, City, State 155 S Madison St, Ste 226, Denver, CO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5407	<p>PROCEDURE MANUAL CFR(s): 493.1251(d)</p> <p>Procedures and changes in procedures must be approved, signed, and dated by the current laboratory director before use.</p> <p>This STANDARD is not met as evidenced by: Based on review of the test procedures manual and staff confirmation, the laboratory director failed to review, sign and date their approval of the procedures for quality assurance, potassium hydroxide (KOH) examination of skin, scabies preparations, histopathology tissue biopsies and Mohs surgery since assuming the directorship in November 2020.</p>