

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  06D0515585	<b>(X3) Date Survey Completed</b>  09/13/2023
<b>Name of Provider or Supplier</b>  Middle Park Health - Kremmling	<b>Street Address, City, State</b>  214 S 4th St, Kremmling, CO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2016</b>	<p><b>SUCCESSFUL PARTICIPATION</b> CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on a routine desk review of the CMS-155 report for Proficiency Testing (PT) performance and email communications with the laboratory director, the laboratory failed to achieve satisfactory performance scores for the American Proficiency Institute (API) PT for Virology testing for two out of three PT events, (event 3 in 2022 and event 2 in 2023). See D2056.</p>
<b>D2056</b>	<p><b>VIROLOGY</b> CFR(s): 493.831(a)</p>

Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.

This STANDARD is not met as evidenced by:

Based on a routine desk review of the CMS-155 report for Proficiency Testing (PT) performance and email communications with the laboratory director, the laboratory failed to achieve a score of 80% for the American Proficiency Institute (API) PT for Virology testing for event 3 in 2022, and event 2 in 2023. Findings include: 1. A review of the CMS-155 Individual Laboratory Profile on 9/13/2023 at approximately 1:15 PM, revealed the API PT for Virology testing scores was 70% for event 3 in 2022 and was 70% for event 2 in 2023. 2. Email communication with the laboratory director on 9/22/2023 at 12:04 PM confirmed the unsuccessful PT scores for Virology.