

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  06D0516905	<b>(X3) Date Survey Completed</b>  12/17/2025
<b>Name of Provider or Supplier</b>  Yuma District Hospital	<b>Street Address, City, State</b>  1000 W 8th Ave, Yuma, CO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	Based on an on-site recertification survey conducted on December 17, 2025, deficiencies were cited for Yuma District Hospital in Yuma, Colorado.
<b>D2000</b>	<p>ENROLLMENT AND TESTING OF SAMPLES CFR(s): 493.801</p> <p>Each laboratory must enroll in a proficiency testing (PT) program that meets the criteria in subpart I of this part and is approved by HHS. The laboratory must enroll in an approved program or programs for each of the specialties and subspecialties for which it seeks certification. The laboratory must test the samples in the same manner as patients' specimens. For laboratories subject to 42 CFR part 493 published on March 14, 1990 (55 FR 9538) prior to September 1, 1992, the rules of this subpart are effective on September 1, 1992. For all other laboratories, the rules of this subpart are effective January 1, 1994.</p> <p>This CONDITION is not met as evidenced by: Based on the surveyor review of 2024 and 2025 Proficiency Testing (PT) documents, surveyor review of testing performed, and an interview with the technical consultant, the laboratory failed to enroll in a PT program for blood culture since 2024. Findings include: 1. Review of 2024 and 2025 PT documents revealed the laboratory was not enrolled in PT for blood culture testing. 2. Review of the laboratory test list revealed the laboratory performed blood culture testing since 2024. 3. An interview with technical supervisor on December 17, 2025, at 2:16 PM confirmed that the laboratory reports growth or no growth on blood cultures, but failed to enroll in blood culture PT since 2024.</p>