

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 06D0518574	(X3) Date Survey Completed 06/17/2026
Name of Provider or Supplier San Luis Valley Health Conejos County Hospital	Street Address, City, State 19021 Us Hwy 285, La Jara, CO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	Based on an on-site initial certification survey conducted on June 17,2026, deficiencies were cited for San Luis Valley Health Conejos County Hospital in La Jara, Colorado.
D5775	COMPARISON OF TEST RESULTS CFR(s): 493.1281(a)(c) (a) If a laboratory performs the same test using different methodologies or instruments, or performs the same test at multiple testing sites, the laboratory must have a system that twice a year evaluates and defines the relationship between test results using the different methodologies, instruments, or testing sites. This STANDARD is not met as evidenced by: Based on a review of laboratory instruments comparison documents, and an interview with the General Supervisor 2 (GS2) according to the Centers for Medicare and Medicaid Services (CMS) Form-209, the laboratory failed to perform biannual evaluations to define the relationship between test results using automated and manual differential. There were approximately 4698 automated differential and 4587 manual differential performed annually. Findings: 1. On June 17, 2026, at approximately 1:30 p.m., a review of the laboratory instrument comparison documents revealed that the laboratory did not perform biannual evaluations to compare automated and manual differential in 2024, 2025, and 2026. 2. On June 17, 2026, at approximately 2:00 p.m., a interview with GS2 confirmed that the laboratory did not perform biannual evaluations to compare automated and manual differential in 2024, 2025, and 2026.