

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  06D0519602	<b>(X3) Date Survey Completed</b>  06/03/2025
<b>Name of Provider or Supplier</b>  Rangely District Hospital	<b>Street Address, City, State</b>  225 Eagle Crest Dr, Rangely, CO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	Based on an on-site recertification survey conducted on June 03, 2025, deficiencies were cited for Rangely District Hospital in Rangely, Colorado.
<b>D5215</b>	<p><b>EVALUATION OF PROFICIENCY TESTING PERFORMANCE</b> CFR(s): 493.1236(b)(2)</p> <p>The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results).</p> <p>This STANDARD is not met as evidenced by: Based on a review of the laboratory's policies and procedures manual, proficiency testing (PT) records review, and an interview with general supervisor #2 (GS2), the laboratory failed to establish a written policy or procedure for, and failed to evaluate PT results that were not evaluated or scored by the PT provider since the laboratory's last survey on February 27, 2024. The laboratory performs approximately 63,335 tests annually. Findings include: 1. A review of the laboratory's policies and procedures manual revealed the laboratory failed to establish a written policy or procedure for evaluating PT scores that were not evaluated or scored by the PT provider since the last survey was conducted on February 27, 2024. 2. A review of the laboratory's PT records revealed the laboratory did not evaluate the accuracy of any analyte for which the PT provider did not evaluate or score since the last survey was conducted on February 27, 2024. 3. An interview with GS2 on June 03, 2025, at approximately 11:00 AM, confirmed that the laboratory failed to establish a written policy or procedure for, and evaluate any PT scores that the PT provider did not evaluate or score since the laboratory's last survey was conducted on February 27, 2024.</p>