

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 06D0879326	(X3) Date Survey Completed 03/09/2018
Name of Provider or Supplier Advanced Dermatology Of Colorado	Street Address, City, State 1006 Centre Ave, Fort Collins, CO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: . Based on the lack of accuracy verification (comparison) testing documentation and staff interview, the laboratory failed in 2016 and 2017 to twice annually verify the accuracy of the scabies skin preparation examination test. Staff confirmed no scabies comparison testing was conducted in 2016 and 2017. Approximately 2 patient specimens are tested annually.</p>