

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 06D0965887	(X3) Date Survey Completed 05/17/2019
Name of Provider or Supplier Dermatology Center Of Steamboat Springs	Street Address, City, State 940 Central Park Dr, Ste 210, Steamboat Springs, CO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5473	<p>CONTROL PROCEDURES CFR(s): 493.1256(e)(2)(g)</p> <p>(e) For reagent, media, and supply checks, the laboratory must do the following: (e) (2) Each day of use (unless otherwise specified in this subpart), test staining materials for intended reactivity to ensure predictable staining characteristics. Control materials for both positive and negative reactivity must be included, as appropriate. (g) The laboratory must document all control procedures performed.</p> <p>This STANDARD is not met as evidenced by: Based on the lack of quality control records and staff confirmation, the laboratory failed to document the quality of the hematoxylin and eosin stain on biopsy slides for intended reactivity each day of use in 2018 and 2019 and approximately 3000 histopathology slides stained annually.</p>