

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 06D2036649	(X3) Date Survey Completed 01/31/2018
Name of Provider or Supplier Peak Dermatology Professional, Llc	Street Address, City, State 2009 W Littleton Blvd, Littleton, CO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5473	<p>CONTROL PROCEDURES CFR(s): 493.1256(e)(2)(g)</p> <p>(e) For reagent, media, and supply checks, the laboratory must do the following: (e) (2) Each day of use (unless otherwise specified in this subpart), test staining materials for intended reactivity to ensure predictable staining characteristics. Control materials for both positive and negative reactivity must be included, as appropriate. (g) The laboratory must document all control procedures performed.</p> <p>This STANDARD is not met as evidenced by: Based on the lack of quality control records and staff confirmation, the laboratory failed to document the staining quality of the hematoxylin and eosin stain (H&E) on biopsy slides for intended reactivity each day of use in 2016 and 2017 and approximately 1000 histopathology slides are stained annually.</p>