

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 06D2117690	(X3) Date Survey Completed 10/12/2020
Name of Provider or Supplier Colorado Center For Dermatology, Pllc	Street Address, City, State 916 S Main St, Ste 201, Longmont, CO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1) At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. This STANDARD is not met as evidenced by: Based on the lack of accuracy verification (comparison) testing documentation and staff interview, the lab failed in 2019 to twice annually verify the accuracy of the Potassium Hydroxide (KOH) skin preparation examination test. Staff confirmed that there were 14 patient specimens tested for KOH in 2019.