

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 06D2124696	(X3) Date Survey Completed 03/13/2025
Name of Provider or Supplier Middle Park Health - Winter Park	Street Address, City, State 47 Cooper Creek Way, Ste 220, Winter Park, CO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The following deficiencies are a result of a desk review of proficiency testing scores obtained from the national database and verified with the proficiency testing provider. The facility was found to be out of compliance with the conditions of the CLIA program. The following condition level deficiencies were found to be out of compliance: 42 C.F.R. 493.803 Condition: Successful Participation [proficiency testing];
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on a routine desk review of the CMS-155 report for proficiency testing performance and review of proficiency testing evaluation reports from the proficiency</p>

testing provider, American Proficiency Institute, the laboratory failed to achieve satisfactory performance scores for two consecutive events in the American Proficiency Institute proficiency testing for the analyte: pO₂, Blood Gas, in event 3 in 2024 and event 1 in 2025, see D2096.

D2096

ROUTINE CHEMISTRY
CFR(s): 493.841(f)

(f) Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:
Based on a routine desk review of the CMS-155 report for Proficiency Testing (PT), and review of evaluation reports provided by the PT provider, American Proficiency Institute (API), the laboratory failed to achieve a satisfactory PT score for two consecutive events for the analyte: pO₂, Blood Gas for event 2 in 2024 and event 3 in 2024. Findings include: 1. A review of the CMS-155 Individual Laboratory Profile on March 13, 2025, at 10:30 AM, revealed the API PT results for pO₂, Blood Gas testing scores for event 3 in 2024 was 0%, and 60% for event 1 in 2025. 2. A review of the PT performance summary provided by API, confirmed the laboratory failed to achieve satisfactory testing scores for two consecutive events for the analyte: pO₂, Blood Gas. Scores were 0% in event 3 of 2024, and 60% for event 1 in 2025.