

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  06D2156283	<b>(X3) Date Survey Completed</b>  11/22/2024
<b>Name of Provider or Supplier</b>  Afc Urgent Care Castle Rock	<b>Street Address, City, State</b>  5700 New Abbey Lane, Suite D-300, Castle Rock, CO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2016</b>	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on a routine desk review of the CMS-155 report for proficiency testing performance and review of proficiency testing evaluation reports from the proficiency testing provider, AAB-MLE, the laboratory failed to achieve satisfactory performance scores for the AAB-MLE proficiency testing for HCT (Non-waived) for two consecutive proficiency testing events, (event 1 in 2024 and event 2 in 2024), see D2130.</p>
<b>D2130</b>	<b>HEMATOLOGY</b>

CFR(s): 493.851(f)

Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on a routine desk review of the CMS-155 report for Proficiency Testing (PT), and review of PT evaluation reports from the PT provider, AAB-MLE, the laboratory failed to achieve a satisfactory score for HCT (Non-waived) for event 1 in 2024 and event 2 in 2024. Findings include: 1. A review of the CMS-155 Individual Laboratory Profile on September 19, 2024, at approximately 11:02 AM, revealed the AAB-MLE PT results for HCT (Non-waived) testing scores for event 1 in 2024 was 60%, and 0% for event 2 in 2024. 2. A review of the PT evaluation reports from the PT provider, AAB-MLE, on November 22, 2024, at 1:24 PM, confirmed that the laboratory failed to achieve satisfactory PT scores for HCT (Non-waived for event 1 in 2024 and event 2 in 2024.