

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 06D2191628	(X3) Date Survey Completed 10/11/2022
Name of Provider or Supplier Urgent Care Lakewood	Street Address, City, State 2616 E 120th Ave, Thornton, CO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on a routine desk review of the CMS-155 report for Proficiency Testing (PT) performance and email communications with the laboratory director of operations, the laboratory failed to achieve satisfactory performance scores for the American Proficiency Institute (API) PT for Cell I.D or WBC Diff testing for two out of three PT events in 2022 (events 1 and 2). See D2130.
D2130	HEMATOLOGY CFR(s): 493.851(f)

Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on a routine desk review of the CMS-155 report for Proficiency Testing (PT) performance and email communications with the laboratory director of operations, the laboratory failed to achieve a score of 80% for the American Proficiency Institute (API) PT for Cell I.D. or WBC Diff for events 1 and 2 in 2022. Findings: 1. A review of the CMS-155 Individual Laboratory Profile on 10/11/2022, at 12:30 pm, revealed the API PT for Cell I.D. or WBC Diff testing scores for 2022 event 1 was 47% and for event 2 was 7%. 2. Email received from the laboratory director of operations on 10/13/2022, at 11:32 am, confirmed two consecutive unsuccessful PT scores for Cell I.D. or WBC Diff testing due to clerical errors.