

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 06D2201195	(X3) Date Survey Completed 01/04/2024
Name of Provider or Supplier Kfhp Premier Medical Offices Laboratory	Street Address, City, State 3920 N Union Blvd, Colorado Springs, CO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on a record review and an interview with the technical consultant (TC) and the quality manager, the laboratory failed to establish and follow a policy to assess two out of two technical consultant competencies since the last survey on 06/11/2021. Findings include: 1. A review of the laboratory's competency records revealed that the laboratory failed to establish and document two out of two TC competencies. 2. An interview with the TC and the quality manager on 01/04/2024, at 1:30 p.m., confirmed that the TC competencies were not established or performed.</p>