

| | | |
|--|--|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 07D0696809 | (X3) Date Survey Completed 09/05/2018 |
| Name of Provider or Supplier Quest Diagnostics | Street Address, City, State 7 Germantown Rd, Danbury, CT | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|---|
| D5417 | <p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor observation and staff interview the laboratory failed to use in date vacutainers for blood collection. Findings include: 1. Surveyor observation of the laboratory's blood collection facility supplies on 9/5/18 at 10:30 AM revealed the following vacutainers used for blood collection were expired. a. Six BD vacutainer dark blue top tubes (Lot# 7180594) with an expiration date of 7/31/18. b. Four BD vacutainer gray top tubes (Lot# 7100990) with an expiration date of 8/31/18. 2. Staff interview with the technical consultant on 9/5/18 at 10:30 AM confirmed the above findings.</p> |